Daily Measurements

Take one measurement each day unless advised otherwise by your healthcare provider. The LED will turn RED each morning and then turn GREEN once a measurement has been successfully transmitted.





Measurement Needed

Measurement Successful

For further assistance, please visit Tenovi.com/support.

tenoví

© 2025 Tenovi Co.

All rights reserved. Republication, duplication or modification in whole or in part is prohibited without prior written permission from Tenovi. All trademarks are owned by or licensed to Tenovi, Co. unless otherwise noted.

Revision 05.01.2024


Power Your Gateway

Plug your Gateway into an outlet with the provided power cord and wait until LED on the Gateway turns red.



Insert BPM Batteries

Insert the four batteries into the back of your blood pressure monitor and connect the cuff into the port on the left side of the monitor.





(4AA Batteries included)

2

Pairing Your BPM & Gateway


Press and hold the "START" button until "Pr" is displayed. The gateway should flash yellow and beep after 5 seconds. Your device will display "Set" when completed properly. This pairing process only needs to be completed once.





Fasten The Cuff

Fasten the cuff to your left upper arm directly on your skin. Leave 0.5 inches between the cuff and your elbow joint. The hose should rest over your forearm.



Properly Set Up

Sit comfortably with your left arm resting on the table in front of you and your palm facing up. Do not cross your legs. Sit quietly and take 5-6 deep breaths.



Take A Measurement

Press the "Start/Stop" button to take a measurement. After measurement has been displayed for **5 seconds**, you may remove the cuff. If the LED ring is GREEN, you have properly taken a measurement for the day, and that measurement has been successfully transmitted.



